



XRF Lab Test Report

Customer : Royal Import Export Center Inc - Canada

Supplier of samples: Royal Mineral Milling and Export Center Corp

Date: 26 March 2019

Instrument: Explorer XRF

Ex-factory Production Date: 2/3/2019

Last Calibration Date: 2/5/20189

Detector: SDD Detector

Vacuum System: No

Test Sample number: Copper-1 201903-4

Sample Location: Dallas Texas.

Test conducted by: Skyray Instruments USA Inc

The information below is for customer knowledge only and is not to be used for any legal action.

Result: Customer provided mineral rock samples were collected from the test location. Laboratory XRF Testing was conducted using Handheld XRF System, as a result of XRF inspection- 25.95% Copper (CU) was found in the rock sample.

Below is conducted full XRF Test results

Sample	No.	Working curve	Supplier	Weight	Date	Voltage	Lines	Mg	Al	Si	P	S	K
Copper-1	201903-4	ORE		2.6	3/26/2019 4:44:32 PM	45		1.1359	2.92454	4.49689	0.43021	1.88817	0.08969

Ca	Ti	V	Cr	Mn	Fe	Co	Ni	Cu	Zn	As	Zr	Nb	Mo	Ag	Cd	Sn	Sb	W	Au	Pb	Ba	Hf	Ta	Hg	Bi	Rb
0.0534	0	0	0	0	11.795	0.2877	0	25.95053	0.05583	0	0.00652	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Best Regards

Name: Stan Brenner
Date: 3/28/2019